

Certificate of Conformance

Product Name: BioDuo Pro 18L

Catalogue No.: RDACC159-M

Serial No.:

Mfg. Date: [01st September 2025]

Test	Results
Appearance	Pass
H/W Version	3.0.0
S/W Version	
RPM Accuracy ¹	Pass
Flow Rate Accuracy ²	Pass
Dosing Accuracy ³	Pass
Communication	Pass

This product with [Serial Number] & RDACC159-M is a part of Resolve Biotech's Quality Excellence Program. In addition to rigorous QC testing, we are committed to ensuring customer satisfaction, with all service matters addressed on priority.

Quality Department

Sales & Service Department

Note: The BioDuo peristaltic pump is tested using platinum cured silicone tube with water at ambient temperature at Zero level of suction and discharge.

¹ RPM Accuracy was tested & certified at Minimum, Medium & Maximum RPM at varied time intervals for 1 hour run.

² Flow Rate Accuracy was tested & certified at Minimum & Maximum Flow Rate at varied time intervals for 8 hours run.

³ Dosing Accuracy was tested by dosing 100 mL for 25 cycles.